

Maine Immunization Program Refrigerator Temperature Log (Fahrenheit)

PIN #

Month /

Page 1

Fill in the circle to reflect your Refrigerator temperature

Day		Time	Office Closed	Too Cold		In Range, But Close to Limit			Target Area Ideal Range							In Range, But Close to Limit				Too Warm	MIN Temp 24 Hours	MAX Temp 24 Hours	Initials
						35	36	37	38	39	40	41	42	43	44	45	46						
1	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
2	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
3	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
4	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
5	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
6	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
7	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
8	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
9	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
10	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
11	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
12	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
13	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
14	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
15	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
16	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						
17	OPEN		☐			☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐					
	CLOSE					☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐						

REPORT AND CORRECT ANY TEMPERATURE BELOW 35 DEGREES F IMMEDIATELY! CALL THE MANUFACTURER TO CHECK VIABILITY

REPORT AND CORRECT ANY TEMPERATURE ABOVE 46 DEGREES F IMMEDIATELY! CALL THE MANUFACTURER TO CHECK VIABILITY